



नेपाल सरकार
स्वास्थ्य मन्त्रालय
स्वास्थ्य सेवा विभाग

राष्ट्रिय जनस्वास्थ्य प्रयोगशाला

टेकु, काठमाडौं

नेपाल सरकार
स्वास्थ्य मन्त्रालय
राष्ट्रिय जनस्वास्थ्य प्रयोगशाला
टेकु, काठमाडौं
२०७८

प.सं.: २०७४।०७५

मिति :- २०७४/०७/१६

बिषय :- संशोधन गरिएको सूचना ।

यस प्रयोगशालामा मिति २०७४।०७।१२ मा बोलपत्र सम्बन्धि निर्देशनको द.१ मा व्यवस्था भए अनुसार भएको बोलपत्र पेश गर्नु पूर्व गरिएको बैठकको निर्णयानुसार बोलपत्र सूचना नं. NPHL/ME/NG/NCB/4-5/2074/075 बोलपत्र नं. ४ र ५ मा रहेको तपसिल अनुसारको मेशिनरी उपकरणको आइटममा भएको स्पेशिफिकेशनमा तपसिल अनुसार कायम गर्ने गरी संशोधन गरिएको व्यहोरा सम्बन्धित सबैमा अनुरोध छ ।

बोलपत्र नं. ४, स्लाइस नं. १, आइटम नं. १: Autoclave A (30 L)

क्र.सं	भएको	हुनुपर्ने
4.2	Made of stainless steel	Made of stainless steel or aluminium

बोलपत्र नं. ४, स्लाइस नं. १, आइटम नं. २: Autoclave B (40 L)

क्र.सं	भएको	हुनुपर्ने
4.2	Made of stainless steel	Made of stainless steel or aluminium

बोलपत्र नं. ४, स्लाइस नं. ४ : Blood Bank Refrigerated Centrifuge

क्र.सं	भएको	हुनुपर्ने
4.7	Speed: Upto 5000 RPM or more, with blood bags (adjustable).	Speed: Upto 4500 RPM or more, with blood bags (adjustable).
4.13	Program memory: Minimum 50.	Program memory: Minimum 10.

बोलपत्र नं. ४, स्लाइस नं. ७, आइटम नं. २ : Fully Automated Biochemistry Analyzer

क्र.सं	भएको	हुनुपर्ने
4.17	External Branded PC with configuration: CPU Pentium core i3, 2.7 GHz or above; minimum 4 GB RAM; minimum 500 GB Hard Disk Drive; High Speed DVD ROM; Keyboard (IOS), Mouse and Mouse Pad; Preloaded latest MS Windows Versions activated; Monitor size 19" or higher; latest anti-virus with one year validity.	Inbuilt or external Branded PC with configuration: CPU Pentium core i3, 2.7 GHz or above; minimum 4 GB RAM; minimum 500 GB Hard Disk Drive; High Speed DVD ROM; Keyboard (IOS), Mouse and Mouse Pad; Preloaded latest MS Windows Versions activated; Monitor size 19" or higher; latest anti-virus with one year validity.

बोलपत्र नं. ४, स्लाइस नं. ७, आइटम नं. ३ : Semi Automated Biochemistry Analyzer

क्र.सं	भएको	हुनुपर्ने
4.10	Heating Block of minimum 12 positions should be available	Inbuilt or external heating Block of minimum 12 positions should be available

(डा.राज कुमार महतो)
निर्देशक